



Test Report

No.: T32020270243SN

Date: JUL 08, 2020

Page 1 of 2

GOLDPRO MEDICAL LIMITED
FLAT A2, 3/F., PAHSANG INDUSTRIAL BUILDING, NO. 16 SAN ON STREET, TUEN MUN, HONG KONG

The content of this test report is extracted from the test report number T32020250103SN where the sample is claimed to be identical.

The following samples were submitted and identified on behalf of the client as:

SURGICAL MASK

SGS Case No. : CA320202504509
 Lot No. / Batch Code : NOT PROVIDED
 Sample Description : BLUE MASK
 Quantity Submitted : 100 PCS
 Sample Receiving Date : MAY 06, 2020
 Test Performing Date : MAY 06, 2020 TO JUN 17, 2020

Test Requested : Please refer to the result summary.

Test Method & Results : Please refer to next page(s).

Result Summary :

Test Requested	Result Summary
Bacterial filtration efficiency (BFE) (EN14683:2019+AC:2019 Appendix B)	See Result
Determination of Breathability (EN14683:2019+AC:2019 Appendix C Differential pressure)	See Result

Signed for and on behalf of
SGS Hong Kong Ltd.

Au Kam Chi, Gigi
Technical Manager

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Test Report

No.: T32020270243SN

Date: JUL 08, 2020

Page 2 of 2

Bacterial filtration efficiency (BFE) (EN14683:2019+AC:2019 Appendix B)

Test Side : White Color (Inside)
 Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.
 BFE Flow Rate : 28.3 l/min
 Test bacteria : Staphylococcus aureus ATCC 6538

Test Specimen	Percent BFE (%)
1	99.7
2	99.9
3	99.5
4	99.5
5	99.7


Determination of Breathability (EN14683:2019+AC:2019 Appendix C Differential pressure)

Test Side : White Color (Inside)
 Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.
 Test Area : 4.9 cm²
 Flow Rate : 8 l/min

Test Specimen	ΔP (mm H ₂ O/cm ²)	ΔP (Pa/cm ²)
1	2.4	24
2	2.7	26
3	2.5	25
4	2.7	26
5	2.7	26

Remark: The test results contained in this Test Report was conducted by a SGS assessed competent subcontractor laboratory

Photo Appendix



This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.