



# Test Report

No.: T32020260563SN

Date: JUL 09, 2020

Page 1 of 3

GOLDPRO MEDICAL LIMITED  
FLAT A2, 3/F., PAHSANG INDUSTRIAL BUILDING, NO. 16 SAN ON STREET, TUEN MUN, HONG KONG

The following samples were submitted and identified on behalf of the client as:

## SURGICAL MASK

SGS Case No. : CA320202629459  
Lot No. / Batch Code : NOT PROVIDED  
Sample Description : BLUE MASK  
Quantity Submitted : 100 PCS  
Sample Receiving Date : JUN 22, 2020  
Test Performing Date : JUN 22, 2020 TO JUL 09, 2020

Test Requested : Please refer to the result summary.

Test Method & Results : Please refer to next page(s).

Result Summary :

Test Requested	Result Summary
Microbial cleanliness (Bioburden) (EN 14683:2019+AC:2019 Annex D)	See Result

\*\*\*\*\*

Signed for and on behalf of  
SGS Hong Kong Ltd.

Au Kam Chi, Gigi  
Technical Manager

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



**Microbial cleanliness (Bioburden) (EN 14683:2019+AC:2019 Annex D)**

**Test Methods**

Bioburden

The analyses were performed according to EN 14683:2019+AC:2019 Annex D and ISO 11737-1:2018

**Test Results**

SURGICAL MASK			
SGS Sample No.:HKHC200600002220-101			
Article Number	Mask Weight	Total Bioburden, cfu/mask	Total Bioburden, cfu/g
1	3.36g	6	1.79
2	3.29g	6	1.82
3	3.36g	15	4.46
4	3.38g	<3	<0.89
5	3.33g	<3	<0.90
Mean:		<6.6	<2.0

Recovery Efficiency	Correction Factor
99.4%	1.0

**Microbial Cleanliness (Bioburden): <2.0 cfu/g**

Standard requirement#: ≤30 cfu/g


**Note:**

1. Results reported on the submitted sample on an as received basis.
2. < = less than
3. cfu = Colony Forming Units
4. Extraction method: by stomacher at 250rpm for 5 minutes
5. # EN 14683:2019+AC:2019 - Medical face masks - Requirements and test methods – Performance requirements for medical face masks – Microbial cleanliness

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

**Photo Appendix**



\*\*\* End of Report \*\*\*

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.